



V. B. S. Purvanchal University, Jaunpur-222003, UP

E-Tender Notice No: PU-ET 36/2020

Centre for Renewable Energy, Prof. Rajendra Singh (Rajju Bhaiya) Institute of Physical Sciences for Study and Research

E-Tender notice offers are invited for the supply and installation of laboratory equipment “Modular solar cell Testing Platform” (**Annexure I**) in Centre for Renewable Energy, Prof. Rajendra Singh (Rajju Bhaiya) Institute of Physical Sciences for Study and Research from manufacturers / authorized dealers / distributors / Channel Partners. Tender shall accompany tender fee of Rs. 1000/- (One Thousand Rupees Only) as tender form cost non-refundable and earnest money deposit (EMD) of Rs. 65,000 /- (Sixty Five Thousand Rupees Only) for **Annexure I**. Tender Fee and EMD should be remitted in the form of RTGS/NEFT in favour of “**Finance Officer**”, **Veer Bahadur Singh Purvanchal University, Jaunpur** (A/C 4539000100038627 IFSC – PUNB0453900) from any of the public sector or private bank. Exemption of EMD is permissible if applicable as per rules. The EMD will be valid up to 45 days beyond the final bid validity period. EMD of the unsuccessful bidder will be returned to them at the earliest after expiry of the final bid validity and latest on or before the 30 days after the award of the contract.

Tender forms along with specifications and other details can be purchased online from the Portal <https://etender.up.nic.in> during 13/06/2020 (document download start date) to 20/06/2020 (online bid submission end date) time 5:00 PM.

Registrar

Notice Inviting Tender (NIT)
For
Supply, installation and commissioning of laboratory equipment
“Modular solar cell Testing Platform”
in
Centre for Renewable Energy, Prof. Rajendra Singh (Rajju Bhaiya) Institute of Physical
Sciences for Study and Research
at
Veer Bahadur Singh Purvanchal University, Jaunpur-222003, UP

E-Tender Notice No: PU-ET 36/2020

CRITICAL DATES SHEET

S. No.	Particulars	Date	Time
1.	Publishing Date	13/06/2020	04.00 PM
2.	Document Download / Sale Start Date	13/06/2020	04.00 PM
3.	Seek Clarification Start Date	14/06/2020	10.00 AM
4.	Seek Clarification End Date	19/06/2020	05.00 PM
5.	Bid Submission Start Date	14/06/2020	10.00 PM
6.	Bid Submission Closing Date	20/06/2020	05.00 PM
7.	Bid Opening Date	22/06/2020	11.00 AM

**Equipment Name for Centre for Renewable Energy, Prof. Rajendra Singh
(Rajju Bhaiya) Institute of Physical Sciences for Study and Research**

S.No.	Name of the Equipment	Qty	EMD (₹)
1.	Modular solar cell Testing Platform	01	

Terms & Conditions

For the tender regarding the supply of “Modular solar cell Testing Platform” for Centre for Renewable Energy, Prof. Rajendra Singh (Rajju Bhaiya) Institute of Physical Sciences for Study and Research, V.B.S. Purvanchal University, Jaunpur - 222003

Bids are invited for the supply of the following item at Prof. Rajendra Singh (Rajju Bhaiya) Institute of Physical Sciences for Study and Research, VBS Purvanchal University, Jaunpur. The bids will be opened on the scheduled date and time in the office of finance officer, VBS Purvanchal University, Jaunpur before a committee approved by the Honourable Vice-Chancellor. University reserves the right to reject one or all bids without assigning any reason at any stage.

The 'Technical and Financial Bid' should be uploaded only on E-tender website in the specified format. In case of lack of any certificate/registration/required documents, the tender shall be liable to be rejected.

Bid No. 1 Technical Bid

1. Tender should be provided accessories – wise details.
2. Photocopy of Registration certificate of GST No. of the firm should be kept in technical bid.
3. Photo copy of PAN Card of the Firm/Proprietor should be in technical bid.
4. Certificate of Central registration in which manufacturer of the said item should be quoted (In case of the proprietary items only) in technical bid.
5. EMD should be remitted in the form of RTGS/NEFT in favour of “**Finance Officer**”, **Veer Bahadur Singh Purvanchal University, Jaunpur** (A/C 4539000100038627 IFSC – PUNB0453900) from any of the public sector or private bank. The certified copy of EMD should be kept in technical bid.
6. All the rates of the required items/equipment must be quoted including packing charges and installation charges etc.
7. All the rates quoted in the tender will be valid for **one year** from the opening date of tender.
8. Items/equipment required shall be supplied and installed in the lab of Prof. Rajendra Singh (Rajju Bhaiya) Institute of Physical Sciences for Study and Research, Veer Bahadur Singh Purvanchal University, Jaunpur.
9. All the technical literature of the equipment shall be provided at the time of tender. The vender may be required to provide price justification of his quoted rates.
10. The tender will be rejected if it is not submitted in the manner as prescribed in technical bid stated above.
11. All the decisions will be reserved with the Vice-Chancellor of V.B.S. Purvanchal University, Jaunpur.
12. Any litigation will be subject to judicature at Hon’ble Civil Court Jaunpur and Hon’ble High Court Allahabad.
13. **Payment terms**
 - a) 90% payment shall be released after receipt of goods in good order and condition and successful installation, thereafter certificate from the concerned head of the department and balance 10% payment shall be released after submission of performance bank guarantee

(PBG) amounting to 10% of the purchase order value. The PBG will be drawn from any nationalize bank valid for the warranty period plus 60 (sixty) days.

or

In case of imports, payment will be made through Letter of Credit. 90% will be paid after submission of all original shipping documents & balance 10% will be released after satisfactory installation and commissioning along with submission of Performance Bank Guarantee.

- b) ensure mentioning
 - (i) Bank details of the beneficiary and pan number
 - (ii) Full name and address of the beneficiary on whom order has to be placed

14. Clearance and delivery

On arrival of shipment/consignment, the supplier is required to arrange custom clearance and transportation of the consignment up to Prof. Rajendra Singh (Rajju Bhaiya) Institute of Physical Sciences for Study and Research, VBSPU, Jaunpur premises. The Tenderer/ bidder will do all types of clearance work and **formalities to deliver/provide the equipment at the site of installation. The expenses incurred will be paid on actual basis not exceeding the quoted amount.** Institute will provide all types of documentary support including Customs Duty exemption certificate. The custom duty, custom clearance charges, transportation charges etc. will be paid in Indian Rupees. This should be included in your quoted rates.

15. If Vendor is quoting integration equipment from company of different origin; in such case two purchase order can be provided on request.

16. Patent rights: the supplier shall **indemnify** the purchase against all third party claims of infringement of patent, trademark or industrial design rights arising from use of the goods or any part thereof in India.

17. Clubbing of equipment in one envelop is prohibited and such tenders will be rejected.

18. bidders to enclose the following documents: -

- a) current income tax and sales tax clearance certificates and pan no.
- b) banker's solvency certificate
- c) summary of audited statement of accounts for the last three years
- d) customer satisfaction certificate from 05 such organization is to be attached with the technical bid.
- e) name and address of minimum 05 clients to whom such equipment has been supplied in the last Five years should be mentioned.

19. **Liquidated damages:** the stores should be delivered/dispatched to destination and ready for operation not later than the delivery date specified. If the supplier fails to deliver any or all the stores or perform the service by the specified date, liquidated damages at 1% per Month or part thereof in respect of the value of stores will be deducted from the contract price subject to a maximum of 5%. Alternatively, the order will be cancelled and the undelivered stores purchased from elsewhere at the risk and expense of supplier.

20. Warranty / guarantee: this comprehensive onsite warranty / guarantee shall remain valid for 24 months after the goods (or any portion thereof as the case may be) have been delivered and commissioned to the final destination indicated in the contract.

21. **Advanced training of the instrument performance and maintenance must be given to the department's/ centre's representatives after the successfully installation of the instrument. The expenditure of above mentioned training will be borne by the company only.**

22. Only those bidders whose bids have been technically found acceptable will be invited for participation in the price bid.
23. Conditional offer will not be accepted.
24. The institute does not bind itself to offer any explanation to those bidders whose technical bid has not been found acceptable by the evaluation committee of the institute.
25. Country of origin of the goods is to be mentioned.
26. Refund of EMD: The EMD will be returned to unsuccessful Bidder only after the Tenders are finalized. In case of successful Bidder, it will be retained till the successful and complete installation of the equipment.
27. Please Note that the University remains closed during Sundays and all specified University holidays.
28. Fax, E-mail Tender will not be accepted.

IMPORTANT

- i. The vendor/suppliers/participants should respond against each of the tender technical specifications (for every equipment) in a concise point wise manner.**
- ii. The vendor should respond within 48 Hours of any complaint/issues regarding the performance/status of the equipment.**
- iii. Installation, commissioning and training**
 - a. The delivery should be considered complete only after successful commissioning of the instrument/equipments.
 - b. The pre-installation requirements should be communicated to VBS Purvanchal University well in advance of the installation.
 - c. The installation, commissioning and training must be done only by well trained personnel. Training should be provided to at least three scientists/faculty of the Prof. Rajendra Singh (Rajju Bhaiya) Institute. Training should be considered incomplete until each of trainees express their satisfaction.
- iv. Demonstration of all the specified accuracy and repeatability of the instrument.**
- v. VBS Purvanchal University Jaunpur authority may accept or reject any or all the bids in part or in full without assigning any reason and does not bind itself to accept the lowest bid. The University at its discretion may change the quantity / upgrade the criteria/ drop any item or part thereof at any time before placing the Purchase Order.**
- vi. Promptly make arrangements for repair and /or replacement of any damaged item(s) irrespective of settlement of claim.**

Bid No. 2 Financial Bid

1. Financial bid will be opened only when technical bid is found to be qualified.
2. No Conditional financial bid shall be entertained/accepted.

Registrar

Annexure-I

Modular solar cell Testing Platform

S. No.	Parameters / Features	Description
1.	Steady state test	<ul style="list-style-type: none">• IPCE & EQE: white bias light / voltage bias• SUN-JV: (0-2 SUN)• 4-point resistivity: sheet resistance• Stability: MPP tracking
2.	Transient tests	<ul style="list-style-type: none">• Transient J-V• Charge Extraction by Linearly Increasing Voltage (Photo/Dark-CELIV)• Transient Photo-Voltage-Current (TPV-TPC)• Transient Electro-Luminescence
3.	Dynamic tests	<ul style="list-style-type: none">• IMVS-IMPS (Frequency range : 100 mHz – 1 MHz)• Impedance Spectroscopy (IS) (Frequency range : 100 mHz – 1 MHz)
4.	Measurement Stage	<ul style="list-style-type: none">• Thermal control with Peltier element (0 to 100°C)• Vacuum holes for flexible substrate• Up to 10 cm x 10 cm supporting area• Metal mask to define illuminate area up to 1 cm²
5.	High-speed LED	<ul style="list-style-type: none">• Custom wavelength• 1MHz pulse modulation• 0.001 - 2 SUN output power
6.	Photodiode	<ul style="list-style-type: none">• Wavelength Range: 400 - 1100 nm• Active Area: 9.7 mm x 9.7 mm (94 mm²)• Rise/Fall Time 45 ns
7.	LED ring	<ul style="list-style-type: none">• Custom ring of 6 high power LED's• Output power: 0.1 to 200mW/cm² (0.001 to 2 SUNs equivalent)• Spot size: 2 cm² (with variable spot size)• Pulsed or continuous
8.	Monochromatic light Source	<ul style="list-style-type: none">• Tunable Light Source• Lamp Type: Xenon, Ozone Free• Xenon lamp power : 300 Watt• Tunable Range: 300-1800 nm• Spectral Resolution: 1 nm• Interchangeable fixed slits• RS232/GPIB• light intensity control with closed loop monitoring photodiode
9.	High-speed probe	High-speed custom probes shielded with separated force and sense wires for voltage and current measurement. Maximum Current : +/- 100 mA Maximum Voltage : +/- 10 V
10.	4-points probe	Self-leveling 4-point probe for sheet resistance measurements
11.	Calibrated photodiode	UV enhanced silicon photodiode with NIST calibration 250-1200 nm

12.	Standard reference silicon solar cell	National Renewable Energy Laboratory (NREL) calibrated silicon reference cell
13.	LED spectrum	<ul style="list-style-type: none"> • Light stabilization kit with feedback photodiode.
14.	Steady State Portable Solar Simulator	Portable Solar Simulator Class AAA with minimum effective area 40 mm × 40 mm with detachable control panel and built in power supply.
15.	Installation, commissioning and training	<ol style="list-style-type: none"> The delivery should be considered complete only after successful commissioning of the instrument. The pre-installation requirements should be communicated to VBS Purvanchal University well in advance of the installation. The installation, commissioning and training must be done only by well trained personnel. The vendor should provide training to at least 3 persons at the installation site to make them familiar with the smooth operation of the instrument. Demonstrate all the specified accuracy, resolution and repeatability of the measurement on the standard sample.
16.	Guarantee/ Warranty	At least 2 years comprehensive warranty
17.	Spare accessories	All accessories along with the tools should be provided
18.	UPS	Required uninterruptible power source (UPS) with one hour backup should be provided by the supplier.
19.	Computer	Computer with latest version of Windows OS compatible with instruments software. Min System configuration: i7 Processor, 8 GB RAM, 1TB Hard Drive, 23" LED Monitor, DVD Writer, Keyboard & Mouse, Color printer. Upgradable option
20.	All the routines must include advanced post processing software for data fitting and file manager.	